



THE POWER OF **CONNECTED**

January 31, 2018

**URGENT: VOLUNTARY PRODUCT RECALL OF
EYESALINE EYEWASH AND FIRST AID KITS**

Dear Valued Customer,

Reason for Recall: Honeywell has learned of potential for leaks affecting a limited number of lots of Eyesaline eyewash bottles. While these products are filled aseptically, the potential for a leak results in a low risk of contamination of the eyewash solution and possible damage to surrounding materials when used in First Aid Kits. The root cause has been addressed and we are confident that the issue has been resolved in our production process. Safety is our priority, so we are initiating a voluntary recall of affected products distributors have in inventory, as well as any sold to your customers and end-users.

Risk to Health: Although Honeywell has not received and is not aware of any reports of adverse health events related to this issue, exposure to infectious agents due to a compromised container barrier could result in infection and may require treatment with antibiotics.

Instructions to Customers:

The lots in question shipped from our distribution center from October 2014 to August 2017. To simplify the recall process, we have partnered with Stericycle. Please take the following actions to help us remove these items from usage:

1. Identify if the product in your possession is affected by locating the lot number on the eyewash bottles or first aid kits and cross-referencing with the List of Affected Lots and Item Numbers in Appendix A. Reference images 1 through 6 for assistance in locating the lot number.
2. Immediately segregate and stop sales or stop usage of any affected products.
3. If you have further distributed this product, please notify your customers of this recall at once. This document should be posted in your branch and can also be used as the communication to your customers. Any entity in possession of affected lots, can obtain replacement eyewash bottles or full first aid kits directly from Honeywell by following steps 4 to 6 below. If you distributed this product, please also complete steps 7 and 8 below.
4. Create a record of affected lots and quantities in your possession.
5. Contact 855.215.5028 to obtain a prepaid return label.
6. A copy of the record you created in step 4, along with the Return Response Form must be included with your return.
7. Email list of customers you notified to Honeywell3787@stericycle.com. Honeywell will use this information strictly to track responses and notify you regarding response rates.
8. Please retain any undelivered notices for future reference.

We are conducting this recall in cooperation with Health Canada and the U.S. Food and Drug Administration. We sincerely appreciate your response by February 28th, 2018.

If you have any questions, please do not hesitate to contact Customer Service at 855.215.5028, Monday – Friday, 8:00 AM – 5:00 PM EST.

Sincerely,

Isabel Ferreira

Product Marketing Manager-Honeywell Safety Products

LOCATING LOT NUMBERS-the below images highlight the most common locations for lot number location. Actual location may vary.

Eyesaline Eyewash Bottles

Case Label	Bottle	Bottle Close-up
 <p>32-000455-0036 EYESALINE[®] EYEWASH Bottle Lang: Eng./Esp./Fr./Ca Lot No: F16091-61 Exp. Date: 03/2019 Unit Size: 32 oz. Weight: 31# No. Per Unit: 12/cs. MADE IN THE USA</p>		 <p>Solución Isotónica Estéril 32</p> <p>LOT F16091-61 EXP. 03/2019</p>
Image 1	Image 2	Image 3

First Aid Kits

Case Label	First Aid Kit Label	Secondary First Aid Kit Label
 <p>FIRST AID PRODUCTS PREMIERS SOINS/PRODUCTOS DE PRIMEROS AUXILIOS PART # / # PRODUIT / # PRODUCTO 019580-4048 SIZE / GRANDEUR / TAMAÑO 16 / UNIT / UNITES / UNIDADES TYPE / TYPE / TIPO KIT / TROUSSE / BOTQUIN QUANTITY / QUANTITE / CANTIDAD 12 EA EXP DATE: MM/YY LOT#: w771783 PACKAGED IN / EMBALLE AU / EMPAQUE EN / MEXICO COMPONENTS OF U.S. AND OTHER COUNTRIES</p>	 <p>1 EA EXP DATE: 12/19 8 83940 29058 4 LOT #: w775352 FMT-237 PACKAGED IN / EMBALLE AU / EMPAQUE EN / MEXICO COMPONENTS OF U.S. AND OTHER COUNTRIES</p>	 <p>w771783</p>
Image 4	Image 5	Image 6

RETURN RESPONSE FORM

Please check ALL appropriate boxes.

- I have read and understand the instructions provided on this letter.
- I have checked my stock and confirm that none of the recalled lots are in my possession.
- I have checked my stock and have segregated inventory as instructed.
- I created a record of the affected lots and quantities in my possession, which I will include with my return.
- I have notified my customers of this recall.
- I will package the products affected appropriately and return immediately.
- I have NOT received any reports of adverse events associated with the recalled product.
- I have received reports of adverse events associated with the recalled product. Please explain: _____

Company Name: _____

Company Address: _____

Name (printed): _____ Date: _____

Signature: _____

PLEASE INCLUDE A COPY OF THIS FORM WITH YOUR RETURN AND EMAIL to Honeywell3787@stericycle.com.

APPENDIX A-List of Affected Lots and Item Numbers page 1 of 5

Item Lot/Batch No.	Item No.	Item Description
0000122273	32-000440-0000	Travel Pack 1&4 oz BFS
0000122709	32-000440-0000	Travel Pack 1&4 oz BFS
0000124971	32-000440-0000	Travel Pack 1&4 oz BFS
0000127498	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
0000127499	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
0000127500	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
0000129666	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F14293-21	32-000498-0043	1OZ BFS 2/BX 120BX/CS
F14293-22	32-000457-0000	1 oz BFS Eyesaline Tril 144/cs
F14293-22	32-000440-0000	Travel Pack 1&4 oz BFS
F14314-41	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F14314-41	32-000454-0000	STERILE EYEWASH BOTTLES-16 OZ.
F14314-41	32-000460-0000	16OZ SINGLE EYEWASH WALL STATION
F14314-41	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F14314-41	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
F14321-41	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F14321-41	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F14321-41	32-000455-0036	32oz BFS Single Packed 12/case
F14321-41	32-000461-0000	EYESALINE SGL 32OZ WS 8/CASE
F15012-31	32-000462-0000	32OZ DOUBLE EYEWASH WALL STATION
F15012-31	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F15012-31	32-000461-0000	EYESALINE SGL 32OZ WS 8/CASE
F15012-31	32-000461-0036	Eyesaline sgl 32oz WS unitized
F15099-21	32-000440-0000	Travel Pack 1&4 oz BFS
F15141-31	32-000452-0000	4oz BFS Trilingual 24/cs
F15141-32	32-000494-0043	4 OZ BFS BULK 36/CS
F15153-21	32-000455-0000	STERILE EYEWASH BOTTLES-32 OZ.
F15153-21	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F15153-21	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F15154-21	32-000462-0000	32OZ DOUBLE EYEWASH WALL STATION
F15154-21	32-000455-0000	STERILE EYEWASH BOTTLES-32 OZ.
F15154-21	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F15154-21	32-000455-0036	32oz BFS Single Packed 12/case
F15154-21	32-000461-0000	EYESALINE SGL 32OZ WS 8/CASE
F15154-21	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F15156-21	32-000460-0000	16OZ SINGLE EYEWASH WALL STATION
F15156-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F15156-21	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
F15156-21	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F15159-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F15159-21	32-000454-0000	STERILE EYEWASH BOTTLES-16 OZ.
F15159-21	32-000454-0036	16oz BFS Single Packed 12/case
F15159-21	32-000460-0036	Eyesaline sgl 16oz WS unitiz
F15159-21	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F15196-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.

APPENDIX A-List of Affected Lots and Item Numbers page 2 of 5

Item Lot/Batch No.	Item No.	Item Description
F15196-21	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
F15196-21	32-000460-0036	Eyesaline sgl 16oz WS unitiz
F15196-21	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F16027-11	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F16027-11	32-000455-0000	STERILE EYEWASH BOTTLES-32 OZ.
F16027-11	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F16091-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F16091-21	32-000460-0000	16OZ SINGLE EYEWASH WALL STATION
F16091-21	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F16091-21	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
F16091-31	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F16091-31	32-000460-0000	16OZ SINGLE EYEWASH WALL STATION
F16091-31	32-000454-0000	STERILE EYEWASH BOTTLES-16 OZ.
F16091-31	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
F16091-31	32-000460-0036	Eyesaline sgl 16oz WS unitiz
F16091-41	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F16091-41	32-000455-0000	STERILE EYEWASH BOTTLES-32 OZ.
F16091-41	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F16091-51	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F16091-71	32-000455-0000-H5	STERILE EYEWASH BOTTLES-32 OZ.
F16091-71	32-000455-0000	STERILE EYEWASH BOTTLES-32 OZ.
F16091-71	32-000462-0000-H5	32OZ DOUBLE EYEWASH WALL STATION
F17025-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F17025-21	32-000465-0000	DOUBLE 16OZ WALL STATION4/CS
F17159-21	32-000454-0000-H5	STERILE EYEWASH BOTTLES-16 OZ.
F17159-21	32-000454-0000	STERILE EYEWASH BOTTLES-16 OZ.
F17159-21	32-000460-0000	16OZ SINGLE EYEWASH WALL STATION
F17159-21	32-000460-0000-H5	16OZ SINGLE EYEWASH WALL STATION
W629896	242020	EMERGENCY EYE WASH STATION
W635979	Z242020	EMERGENCY EYE WASH STATION
W639573	242020	EMERGENCY EYE WASH STATION
W643768	242020	EMERGENCY EYE WASH STATION
W644740	SF00002799	S0004 EYEWASH KIT FAK PL
W647726	242020	EMERGENCY EYE WASH STATION
W649390	SF00003881	TRUCK TRUMA BAG
W651014	242020	EMERGENCY EYE WASH STATION
W663008	242020	EMERGENCY EYE WASH STATION
W665548	242020	EMERGENCY EYE WASH STATION
W667898	64058004	4638 HEAVY INDUSTRIAL STATS
W668933	145225	DELUXE TRAUMA KIT
W669276	242020	EMERGENCY EYE WASH STATION
W670340	242020	EMERGENCY EYE WASH STATION
W671530	64058000	4641 HEAVY INDUSTRIAL STATS
W671820	34180FP	3 SHELF INDUSTRIAL FAK W/PUMP

APPENDIX A-List of Affected Lots and Item Numbers page 3 of 5

Item Lot/Batch No.	Item No.	Item Description
W673045	64058000	4641 HEAVY INDUSTRIAL STATS
W673046	64058001	4640 HEAVY INDUSTRIAL STATS
W673047	64058001	4640 HEAVY INDUSTRIAL STATS
W673095	Z019727-0014L	24U PL GENERAL BURN KIT
W673780	34140FP	2 SHELF INDUSTRIAL FAK W/PUMP
W674138	64058001	4640 HEAVY INDUSTRIAL STATS
W674139	64058004	4638 HEAVY INDUSTRIAL STATS
W675188	242020	EMERGENCY EYE WASH STATION
W675190	64058000	4641 HEAVY INDUSTRIAL STATS
W676192	68203	FILL TRAUMA KIT
W676718	64058001	4640 HEAVY INDUSTRIAL STATS
W676769	SF00003715	5062 FAK FILL/LOGO 180 LINER
W677130	148805	EMERGENCY MEDICAL FANNY PACK
W677131	148805	EMERGENCY MEDICAL FANNY PACK
W677614	SF00003849	FIRST RESPONDER REFILL
W677991	68203	TRAUMA KIT
W678355	34140FP	2 SHELF INDUSTRIAL FAK W/PUMP
W678367	Z019803	39P247 MAJOR TRAUMA KIT
W678652	6824PIZ	S0503 FAK FILL/24U MT LOG
W678655	68P24AG	S0004 FILL/FAK LOGO 24U PL
W678658	SF00003376	S0006 100P CABINET FILL W/DETEC
W678721	34400LFB	4 SHELF IND W/LINR /LOGO
W678927	34400LFB	4 SHELF IND W/LINR LOGO
W678935	68342ARBS	FAK 342 BOTTOM
W678944	SF00003970	S0006 3 SHLF PLANT KIT
W679237	34140LFP	2 SHELF IND W/LINER FAK/W/PUMP
W679239	34180LFC	3 SHELF IND W/LINR & CPR/LOGO
W679240	34400LFB	4 SHELF IND W/LINR & BBK/LOGO
W679246	68140LMAX	S0293 LOGO/140 STAN
W679247	68400LISS	S0700 IND.SAFETY LOGO/STD FILL W/LNR
W679248	68400LMAX	FAK LOGO/400 STAN
W679249	68460SM	S0776 LOGO FILL 460
W679255	018500-4222	4222 SOFT PACK LARGE NORTH
W679262	34140RF	REFILL FOR 34-140 KIT
W679387	64058001	4640 HEAVY INDUSTRIAL STATS
W679389	64058004	4638 FAKHEAVY INDUSTRIAL STATS
W679390	64058004	4638 HEAVY INDUSTRIAL STATS
W679409	34460RF	REFILL FOR 34-0460 TRUCK KIT
W679789	64058000	4641 HEAVY INDUSTRIAL STATS
W679790	64058001	4640 HEAVY INDUSTRIAL STATS
W679797	34180LFC	3 SHELF IND W/LINR & CPR/LOGO
W679802	6824PIZ	S0503 FILL/24U MT LOG
W679803	68600PROAM	S0006 FILL 5 SHELF IND FA
W679805	68P24AG	S0004 FILL/FAK LOGO 24U PL
W679806	68P24AG	S0004 FAK FILL/FAK LOGO 24U PL

APPENDIX A-List of Affected Lots and Item Numbers page 4 of 5

Item Lot/Batch No.	Item No.	Item Description
W679839	64058000RF	FAK REFILL FOR 5 SHELF
W679840	64058001RF	REFILL FOR 4 SHELF
W680070	SF00001322	4715 4 SHELF FAK
W680075	018500-4222	4222 SOFT PACK LARGE NORTH
W680083	64058004RF	FAK REFILL FOR 3 SHELF
W680084	64058006RF	2 SHELF REFILL
W680088	SF00003220	FAK REFILL FOR SF00003198
W680089	SF00003445	REFILL FOR 2SHLF SF3256
W680092	Z019802	DELUX TRAUMA KIT
W680094	Z63157987	LSS EMERGENCY BURN KIT
W680117	Z019721-0010L	150P CABINET NORTH
W680122	Z019720-0009L	100P CABINET NORTH
W680484	013107-4533	4533 MED CABINET F/A
W680485	34140FBP	S0006 2 SHELF IND. FAK & PUMP
W680486	34140LFB	2 SHELF IND W/LINER & FAK
W680503	019721-0010L	150P CABINET NORTH
W680557	Z346100	STANDARD EMERGENCY MEDICAL KIT
W680809	Z34140LFSP	FOOD SERVICE KIT, 2 SHELF
W680812	68400LWS	FAK FILL/FAK 400WL
W680818	019722-0011L	200P CABINET NORTH
W680820	34400RF	REFILL FOR 34-400 KIT
W680821	34460RF	REFILL FOR 34-0460 TRUCK KIT
W680822	SF00003220	REFILL FOR BNSF KT SF00003198
W680972	64058001	4640 HEAVY INDUSTRIAL STATS
W680974	340420FP	TRK/AUTO STL MED WP FAK W/PUMP
W680975	340460F	TRUCK KIT DELUXE FAK LOGO
W681019	148805	EMERGENCY MEDICAL FANNY PACK
W681027	Z019721-0010L	150P CABINET NORTH
W681028	Z019721-0010L	150P CABINET NORTH
W681038	Z019816	FAK ANSI & OSHA COMPLIANCE PK
W681224	344F	4 SHELF INDUSTRIAL CABINET
W681226	68140BAN	S0006 140 IND. FILL/FAK LOGO
W681228	68400LNODEA	S0006 STANDARD 4 SHELF
W681248	018504-4222	4222 SOFT PAK LARGE NORTH W/ MICRO
W681332	Z346100	STANDARD EMERGENCY MEDICAL KIT
W681442	68400LWS	FILL/FAK 400WL
W681451	SF00002926	5064 5SHLF LOGO&FILL W/LINER
W681453	68203	FAK FILL TRAUMA KIT
W681459	64058000RF	REFILL FOR 5 SHELF
W681460	SF00003219	REFILL FOR SF00003199
W681461	SF00003220	REFILL FOR KT SF00003198
W681465	Z019802	DELUX TRAUMA KIT
W681594	6832649	FAK TRIP FILL/LOGO 140 KIT
W681595	Z019721-0010L	150P CABINET NORTH
W681606	Z346100	STANDARD EMERGENCY MEDICAL KIT
W681748	6832649	FAK TRIP FILL/LOGO 140 KIT

APPENDIX A-List of Affected Lots and Item Numbers page 5 of 5

Item Lot/Batch No.	Item No.	Item Description
W681749	64058001	4640 HEAVY INDUSTRIAL STATS
W681757	68400LISS	S0700 IND.SAFETY LOGO/STD FILL W/LNR
W681767	34420RF	REFILL FOR 34-0420 TRUCK KIT
W681803	34400LFB	4 SHELF IND W/LINR & BBK/LOGO
W682068	SF00003220	REFILL FOR KT SF00003198
W682069	64058004RF	REFILL FOR 3 SHELF
W682095	SF00003198	S0005 36UN MT MED FILL
W682466	011007-4330	4330 16 UN PL FAK
W682471	340460F	TRUCK KIT DELUXE FAK LOGO
W682472	34180LFC	3 SHELF IND W/LINR & CPR/LOGO
W682510	SF00003659-H5	36UN MT FAK
W682546	Z019721-0010L	150P CABINET NORTH
W682550	SF00003513	5169 3 SHELF STD FILL
W682728	SF00003256	S0006 2SHLF FAK FILL
W683351	Z019722-0011L	200P CABINET NORTH
All lots	FAKBLYACHT	First Aid Kit-Canada-Yacht